

On-line Table: MR imaging parameters

	T1WI	T2WI	FLAIR	DWI	TOF MRA	CE MRA
Parameter						
Repetition time, ms	633	4483	8802	9000	25	7.0
Echo time, ms	20	127	126	75.6	6.8	2.4
Echo-train length	1	1	1	1	1	1
Flip angle, °	90	90	90	90	20	25
Section thickness, mm	5	5	5	3	1.4	1.4
Intersection gap, mm	1.5	1	1	1	0	1.5
FOV, mm	220 × 220	220 × 220	220 × 220	240 × 240	240 × 240	299 × 299
Matrix	320 × 192	448 × 256	320 × 192	160 × 160	384 × 224	384 × 224
No. of signals acquired	1	2	1	2	1	1
No. of sections	23	24	24	72	125	128

Note:—CE indicates contrast enhanced.

^a A standard dose of 0.1 mmol/kg of body weight of a gadolinium-based contrast agent (Gadovist, Bayer Schering Pharma, Berlin, Germany) was injected intravenously at a rate of 5 mL/s, followed by a 20-mL bolus of saline at a rate of 5 mL/s using a power injector (Optistar; Mallinckrodt, St Louis, Missouri).